

EU Quality Management System Certificate

We hereby certify the company

bredent medical GmbH & Co. KG
Weißenhornstraße 2
89250 Senden
Germany

the introduction and application of a quality management system in accordance with Annex IX, Chapter I and III of Regulation (EU) 2017/745 for conformity assessment.

An audit by mdc has proven that this quality management system meets the following requirements:

Annex IX – Chapter I (Quality Management System)

of Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices.

Surveillance is carried out in accordance with Annex IX, Section 3 of Regulation (EU) 2017/745.

This certificate from mdc medical device certification GmbH (Notified Body 0483) consists of 3 pages. Details about the devices covered as well as further information and conditions are contained on the following pages.

Valid from 2024-08-05
Valid until 2027-10-01

Registration No. D1146000054
Report No. P23-00466-263806

Stuttgart, 2024-08-05



Notified Body



Devices:

Dental Implants and accessories

Intended purpose: The product is intended to restore and/or correct the function, phonetics and aesthetics of existing teeth and/or lost teeth as part of a dental restoration.
These are dental titanium or zirconium implants for insertion into the jaw bone.

Risk class: IIb

Devices for prosthetic dentistry - accessories and others

Risk class: IIa

dental impression devices for prosthetic dentistry

Risk class: IIa

Odontostomatology instruments - others

Risk class: IIa

Odontostomatology instruments - others

Risk class: I (reusable)

Notes:

In the case of class I devices that are reusable surgical instruments the involvement of mdc is limited to the aspects relating to the reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing as well as the related instructions for use.

For the placing on the market of class III and IIb implantable devices (except sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors within the meaning of Regulation (EU) 2017/745, Art. 52 (4), 2nd paragraph and with the exception of custom-made devices of class III), an EU technical documentation assessment certificate is also required.

The certificate is based on the previous certificate
D1146000051 (2023-06-21)

with the following changes to D1146000051:

Supplemented by

- Odontostomatology instruments - others (Class IIa und Ir)
- Dental Implants and accessories: zirconium implants